

# Hybex™ Microsample Incubator USER MANUAL

Cat. #1057-30-0, 1057-30-2



**FOR RESEARCH USE ONLY**  
**Not for Use in Diagnostic Procedures**

## Serial Number

The following serial number identifies the specific instrument you have purchased and must be referenced when requesting service. A copy is affixed to the instrument.

Technical Service: (408) 733-7337, [techserv@scigene.com](mailto:techserv@scigene.com)

## Warranty

SciGene warrants that the heating unit described in this manual shall be free of defects in materials and workmanship for a period of 12 months from date of delivery. This warranty does not cover removable blocks or accessories. In the event of a defect during the warranty period, SciGene's limit of liability will be to provide replacement parts at no charge or, at its sole discretion, replace the product. The foregoing warranty is void in the event the unit was abused or modified or used in a manner inconsistent with its intended purpose. SciGene makes no other warranty, expressed or implied including warranties of merchantability and fitness for a particular purpose. In no event shall SciGene be liable for any direct, indirect, special, incidental or consequential damages or for any damages resulting from loss arising out of or in connection with the sale, use or performance of the product.

## Copyright

Copyright ©2004-2007 SciGene Corporation. All rights reserved. SciGene and Hybex are trademarks of SciGene Corporation, Sunnyvale, CA. All other trademarks used in this manual are the property of their respective owners.

## Table of Contents

<b>I. SAFETY NOTICES .....</b>	<b>3</b>
A. Warnings .....	3
B. Cautions .....	3
<b>II. UNPACKING AND SET UP .....</b>	<b>4</b>
A. Unpacking the Heating Unit .....	4
B. Parts Provided.....	4
C. Installation .....	4
<b>III. USING YOUR HYBEX SYSTEM.....</b>	<b>5</b>
A. Components, Controls and Accessories .....	5
B. Handling Tube Blocks and Waterbath Insert .....	6
C. Using the Temperature Controller.....	6
D. Calibrating the Temperature Controller.....	7
<b>IV. MAINTAINING YOUR HYBEX SYSTEM.....</b>	<b>8</b>
A. Powering Off.....	8
B. Checking and Replacing Fuses .....	8
C. Cleaning .....	8
<b>V. TROUBLESHOOTING.....</b>	<b>9</b>
<b>VI. SPECIFICATIONS.....</b>	<b>9</b>
<b>VII. ORDERING INFORMATION.....</b>	<b>9</b>

## I. SAFETY NOTICES

### A. Warnings

Failure to comply with the following warnings that are affixed to the product can lead to possible personal injury or death.



### B. Cautions

Failure to comply with the following cautionary statement affixed to the product may lead to possible personal injury.



## II. UNPACKING AND SET UP

### A. Unpacking the Heating Unit

The Hybex™ Microsample Incubator heating base is shipped on its side within the interior carton. Remove any packing material around the unit and lift the unit out; taking care not to damage the hinged cover.

Please note that heat blocks and other accessories are packaged separately from the heating base.

Carefully inspect the heating unit for damage.

*If there is evidence of damage, do not discard the shipping materials since they may be needed to return the unit.*

### B. Parts Provided

The following items are included with the Hybex heating base within the re-closable plastic bag:

- Power cord
- User Manual

### C. Installation

Place the unit on a level surface within a few feet of the power source. The unit should be positioned to have clearance along the top so that the lid can be opened completely without interference. Leave at least 3 inches of clearance from the back panel for air circulation.

Insert one Hybex heat block. See **Section VII** for a list of available heat blocks and **Section III** for how to insert and remove blocks.




***Do not operate the system without a heat block inserted into the heating unit !!***

Plug the provided power cord into the back of the unit and then to a properly grounded outlet. Use only the power cord provided. Turn on power to the system using the switch on the front of the unit.

## III. USING YOUR HYBEX SYSTEM

### A. Components, Controls and Accessories

Name	Function
Power Switch	Turns on main power to unit
Temperature Controller	Used to set and observe block temperature
Heated Lid	Heats the sealed air around the block
Temperature Probe	Senses the temperature of the block
Digital Thermometer Jack	For calibrating system temperature
Heat Block (optional)	Heats tubes and plates without cap condensation
Microarray Incubation Chamber (optional)	For array hybridization
Waterbath Insert (optional)	For heating buffers or water
Digital Thermometer (optional)	For calibrating system temperature



## B. Handling Tube Blocks and the Waterbath Insert

Tube heating blocks and the waterbath insert are placed and removed from the heating unit with the aid of two bolts that act as handles. To properly install a block or insert:

1. Thread the two bolts provided several turns into the block.
2. Grasp the ends of the bolts, lift the block, and then carefully lower it into the unit; guiding it over the temperature probe. The opening in the bottom of the block should fit over the probe.
3. Remove the threaded bolts before using the system.



***Do not operate without a block inserted into the heating unit!!***

## C. Using the Temperature Controller

The temperature controller has a single LED read out and four push buttons. The LED displays the set temperature when the SET key is pressed and held. Otherwise the **ACTUAL** block temperature is displayed. The push buttons are used to set the block temperature and, when required, to calibrate the temperature controller.

1. To set the block temperature, hold in the SET button. The previous set temperature will be displayed.
2. While depressing the SET button, press the up or down arrow buttons until the desired set temperature is shown.
3. Release the buttons and the actual block temperature will again be displayed. The unit will now adjust the heat of the block until the new set temperature is attained.

The controller is calibrated at the factory to provide an accurate block temperature when operated between 35 to 99°C.

#### D. Calibrating the Temperature Controller

The temperature controller comes calibrated from the factory to provide accurate block temperatures from 35 to 99°C.

The temperature controller will require calibration only if:

- When checking the block temperature with a calibrated digital thermometer, the block temperature differs by more than one degree (1°C) from the actual temperature shown on the controller display.

An NIST calibrated digital thermometer (sold separately, SciGene Cat. #1051-52-0) is required to calibrate the unit.

Follow these steps to adjust the controller to achieve accurate temperatures.

1. Set the temperature on the controller to 65°C and allow the temperature in the unit to stabilize.
2. Using the cable provided with the digital thermometer (SciGene Cat. #1051-52-0), plug one end into the blue receptacle found on the back panel of the unit and the other into the digital thermometer.
3. Turn on the thermometer using the **ON/OFF** button on the keypad. The temperature of the block will be displayed.
4. On the temperature controller, press the infinity key for three seconds until “OPEN” appears.
5. Press the down arrow four times until “Cal” appears.
6. Press and hold the set key. The existing offset value between the controller and digital thermometer is displayed.
7. Calculate the difference in the temperature shown on the controller and the digital thermometer; e.g. the controller displays 52.5°C and the digital thermometer displays 51.0°C, the difference is 1.5°C.
8. Press and hold the **SET** key and use the up and down arrows to enter the offset value calculated in Step #7. For example, if the controller displays a temperature that is 1.5°C higher than the digital thermometer, adjust the offset value to *minus* 1.5 (-1.5).
9. Press the **Infinity Key** (∞) to exit calibration. The block temperature is now calibrated for the set temperature you selected. Repeat the calibration process starting at Step 1 for the highest temperature for the range of temperatures you will be using.



## IV. MAINTAINING YOUR HYBEX SYSTEM

### A. Powering Off


Turn the power switch to the OFF position and unplug the power cord before performing any service procedure.

### B. Checking and Replacing Fuses

There are two fuses located in a removable fuse block below the power cord receptacle on the back of the unit.

Insert a small, flat blade screwdriver into the tab recess just below the plug receptacle. Push down to release the fuse block. Slide out the fuse holder from the fuse block while holding the retaining tab out of the way. Gently pry out the fuses. A blown fuse appears dark.

Always replace fuse(s) with those of the same amperage and voltage as shown on the label below the fuse block.



### C. Cleaning

Clean the metal surfaces and heated lid using a mild, detergent-based spray cleaner and wipe with a soft cloth. Avoid abrasive cleaners that can scratch surfaces.

The bottom of the chamber is sealed and will retain spills and condensation. Clean the block chamber and temperature probe frequently to avoid build up of residue that can interfere with the heat control system.

## V. TROUBLESHOOTING

Symptom	Cause	Solution
Power switch light does not turn on	Blown fuse(s)	Replace fuse(s) on back of unit, beneath power cord receptacle.

## VI. SPECIFICATIONS

<b>Electrical</b>	
Cat. #1057-30-00	115V AC; 50/60 Hz; 250W
Cat. #1057-30-02	220V AC; 50/60 Hz; 250W
<b>Dimensions</b>	
Outside (cover closed) H x W x D	8 x 6 x 12 inches (12 x 14 x 30 cm)
<b>Weight</b>	
Net (without block)	5.2 lbs (2.4 kg)
Gross (in shipping carton)	4.1 to 6.0 lbs (1.9 to 2.7 kg)
<b>Performance and Controls</b>	
Temperature Range	Ambient +5°C to 99°C
Temperature Regulation	± 0.2°C
Heat up Time	>5°C per minute
Temperature Controller	Digital PID, single loop
Temperature Display	Actual or Set single LED
Digital Thermometer Output	Thermocouple

## VII. ORDERING INFORMATION

Cat. #	Description	UoM
1057-30-0	Hybex Microsample Incubator, heating base, 115V	Each
1057-30-2	Hybex Microsample Incubator, heating base, 220V	Each
1057-31-0	PCR tube block, Holds 96x0.2ml tubes or single plate	Each
1057-37-0	Hybex Microarray Incubation Chamber-Complete with slide racks, assembly tool and handle	Each
1057-33-0	0.6mL tube heat block, holds 60x0.6mL tubes	Each
1057-34-0	1.5mL tube heat block, holds 32x1.5mL microcentrifuge tubes	Each
1057-35-0	Waterbath Insert for Hybex heating unit	Each
1051-52-0	Digital thermometer. Includes cable and NIST certificate	Each