# Push-To-Set<sup>™</sup> Adult Intermittent Suction Unit Digital and Analog (ANSI)

### **DESCRIPTION AND APPLICATION**

The Push-To-Set<sup>™</sup> Intermittent Suction Unit exceeds clinical expectations in medical suction procedures, by offering technology breakthroughs plus the quality and simplicity you are accustomed to receiving from Ohio Medical Corporation's products. It retains the unique, patented unilogic module which provides a reliable, quiet means of cycling the regulator on and off during intermittent suctioning. It also offers dual modes to provide the option of an intermittent mode for nasogastric drainage or C.A.S.S. procedure and a continuous mode for tracheal, pharyngeal or surgical suctioning.

The Push-To-Set<sup>™</sup> Internittent Suction Unit has an ergonomic design that contains two new exclusive safety features that simplify processes, minimize clinical errors and increase product performance. The patented Push-To-Set<sup>™</sup> mechanism automatically establishes a vacuum limit with each vacuum level setting in both continuous and intermittent modes. A unique dual-spring regulator module ensures precision in the critical care range (0-200 mmHg) while also providing unusually fast adjustment (2 turns) up to full wall vacuum for emergency resuscitation procedures.

#### **FEATURES AND BENEFITS**

- EXCLUSIVE! Patented Push-To-Set<sup>™</sup> Improves patient safety May prevent inadvertent unregulated suction Simple, one handed operation
- EXCLUSIVE! Quick-to-Max Faster adjustment (2 turns from zero) to full available wall vacuum for emergency situations
- Unilogic Timing Module Pneumatic, quiet, reliable operation Returns to atmospheric pressure during "OFF" cycle Timing cycles Independently adjustable 3 to 30 seconds "ON" and "OFF"
  - cycles Factory preset at 15 seconds on/8 seconds off (± 3 seconds) Begins in the "ON" mode

#### Flow Rates

(Intermittent) Preset at 8 L/min (ASTM) (Continuous) 0 to 80 L/min (without fittings at full increase)

Large, Easy to Read Gauge

Color coded ranges: Low(green), Medium(yellow), High(orange)

- Analog: Familiar, industry-standard analog design Digital: Unmatched accuracy for greater safety: ± 1% of full scale
- Safety Features

Push-to-Set<sup>™</sup> automatic occlusion design Positive pressure safety relief valve





#### Modular Components

Common components reduce service part inventory Simplifies service procedures Low cost replacement parts and kits

- Durable Design and Construction Few mechanical parts Rugged, shatter-resistant case Corrosion & Lubrication Free Service-free backplate
- Ordering Options

Available in various configurations including adapters for connection to source vacuum. Also available with options for collection devices and overflow safety protection.



Battery (Digital)	: Primary Lithium: 2/3AA, 3/6V, 1/6 Ah
Dimensions:	6.75" H X 2.8" W X 4.8" D (17.15 x 7.11 x 12.19 cm)
Weight:	1 lb 4oz (0.57 Kg) less fittings and adapters
F.O.B. Point:	Gurnee, IL
Standards:	Conforms to ASTM F960
	ninal, subject to change without notice. Please see "Instructions for Use" rent regulated vacuum ranges.
SHIPPING INFORMATION	
Shipping Weight: 1 lb 11 oz (0.77 Kg) less fittings and adapters	
Package Size:	6.5" H x 10.5" W x 8" D (16.51 x 26.67 x 20.32 cm)
Packaging:	One per carton
Terms:	net 30 days
	Dimensions: Weight: F.O.B. Point: Standards: * Specifications are non manual for the most cur SHIPPING INI Shipping Weigh Package Size: Packaging:

## **Parts Configuration**

<u>8</u>	<u> 77XX - XXXX - 9X</u> X
Push-To-Set <sup>™</sup> Series	Fittings (patient connection)
Adapter (vacuum source connection)   No Adapter with 1/8" NPT Female00   Ohio Diamond01   DISS Hand-I-Twist (HIT)02   DISS Nut & Gland03   Schrader	No Fitting 1/8" NPT Female
O.E.S. Twist	Push-To-Set <sup>™</sup> Adult Intermittent Suction Unit   Analog: Digital:   1251 1351 Coconut(Standard) 1751 Red   (Coconut only) 1451 Yellow 1851 Lt. Blue   1551 Blue 1951 Purple

© 2012 Ohio Medical Corporation.

This document contains information that is proprietary and confidential to Ohio Medical Corporation. Use of this information is under license from Ohio Medical Corporation. Any use other than that authorized by Ohio Medical Corporation is prohibited.

Ohio Medical Corporation and the Ohio Medical Corporation logo are registered trademarks and Push-To-Set is a trademark of Ohio Medical Corporation.

U.S. Patent #6,623,463 (PushTo-Set<sup>™</sup> Vacuum System)



Ohio Medical Corporation , 1111 Lakeside Drive - Gurnee, IL 60031USA TollFree: 866-549-6446 or 1+847-855-0800 - Fax: 1+847-855-6218 www.ohiomedical.com Form No. 255442 (Rev.3) 09/2012